

Test results “Top-Quality Non-Cutting” instruments

Ref. No. PD

09/2792

March 14, 2011

ANALYTICAL / TEST REPORT

In Respect of : Medipharchem Nederland B.V
Sample Received from : Representative
Name of the Sample : Top-Quality Non-Cutting Instrument
Type/Grade of Material : AISI-410
Specification against which tested : **Analytical Results Got on Direct Emission Spectrometer**

Results are in percentage:

	Cr	12		
C	0.15		Mn	1
P	0.04		S	0.03
Si	1		Ni	0.6

This analytical/test report is based solely on the particular sample(s) supplied by the party and should not be reproduced in part.

Sampling has not been performed by the xxx and the xxx does not accept responsibility that the sample(s) supplied is/are truly representative sample(s) of any batch or stock or entire production.

While the xxx agrees to take every reasonable precaution to ensure validity of its test results, it assumes no liability for the accuracy thereof beyond the amount of the fee charged for the analysis/test.

The party shall assume full responsibility for the ethical use of the results in the analysis/test reports and the laboratory shall be held free from any and all claims which may result from the use of such data by the party or others.

After completion of the report the sample will be preserved for three month until negotiated otherwise.

Project Director